


# Important Prober Dimensions for Board Holder Construction

## Access and Access 2 Slot Heights (measured from base floor)



Maximum Recommended Holder plate dimensions (allow for proper mounting of cross bar with end slide):

- Access: 18.6" wide by 14.2" deep
- Access 2: 20.4" wide by 23.5" deep

## Access DH Slot Heights (measured from base floor)

